

Alarm X

### File Information

File Created Date: 14/11/21 17:03:01

Note: All Times shown are based on UTC-10:30 and 24-Hour clock [DD/MM/YY HH:MM:SS]

### Device Information

Device Type:	TZ-TempU04	Firmware Version:	3.02
ID:	TZ0421090150		
Start Delay:	0 min	Log Interval:	1 min
Description:	Temperature Data Logger		

### Logging Summary

First Point:	14/11/21 13:12:58	Max:	34.3°C
Stop Time:	14/11/21 17:01:58	Min:	23.7°C
Number of Points:	230	Average:	27.8°C
Trip Length:	00d 03h 49m 00s	MKT:	28.8°C
Stop Mode:	USB Stop	Start Mode:	Button Start

### Marked Events

N/A

Alarm Condition	Alarm Delay	First Point Time	Time of Violations	No. of Violations	Status
Over 30.0°C	00d 00h 00m 00s	14/11/21 13:59:58	00d 01h 18m 00s	79	Alarm
Below 5.0°C	00d 00h 00m 00s	00/00/00 00:00:00	00d 00h 00m 00s	0	OK



